



Technical Specification of ENT Set

S.N.	Purchaser's Specifications	Bidders Remarks in Detail
	ENT Set	
	<i>Manufacturer</i>	
	<i>Brand</i>	
	<i>Type / Model</i>	
	<i>Country of Origin</i>	
1	Description of Function	
1.1	For routine ENT examination of patient.	
2	Technical Specifications	
2.1	Each ENT set must include at least following devices (1 pcs each): <ul style="list-style-type: none"> • Ophthalmoscope with all accessories • Double Lens Otoscope with all accessories • Bent Arm Illuminator • Nasal Speculum • Tongue Blade Holder • Nasal Tongue Depressor • Laryngoscope Mirrors 	
2.2	<p><u>Optahlamoscope:</u></p> <ol style="list-style-type: none"> a. Ophthalmoscope set composed of diagnostic head mounted on a handle. b. Range of lenses not smaller than -35D to +40D, adjustable in steps c. Anti-reflection lens. d. Ability to magnify at least five times. e. Available apertures at a minimum: small, large and semi-circle, fixation star. f. Colour temperature: Cool white in the range 3,500-4,000 K. g. Red-free, blue, and green filters. h. Handle with on/off switch. i. Provided with an LED light source. 	
2.3	<p><u>Otoscope:</u></p> <ol style="list-style-type: none"> a. Handle with a built-in LED light source in a mounted head. b. Including rotatable viewing lens with a magnification between 3 times. c. Colour temperature: Cool white in the range 3000-4000K. d. Including reusable, autoclavable specula. e. Handle contains on/off switch. f. Includes a port to connect an insufflation bulb. 	


 Biomedical Engineer
 NEC NO: 83941

S.N.	Purchaser's Specifications	Bidders Remarks in Detail
	g. Use 2 x type AA Alkaline batteries as a power source.	
2.4	Should be able to operate with alkaline batteries or 6V power adaptor.	
3	Accessories, spares and consumables	
3.1	Battery: 1 set Bag: 1 pc	
4	Operating Environment	
4.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Power Supply, Climate, Temperature, Humidity, etc.	
5	Standards and Safety Requirements	
5.1	Must submit valid ISO and CE/USFDA approved product certificate.	
6	User Training	
6.1	N/A	
7	Warranty	
7.1	Comprehensive warranty for 2 years after acceptance.	
8	Maintenance Service During Warranty Period	
8.1	During warranty period supplier must ensure corrective/breakdown maintenance whenever required.	
9	Installation and Commissioning	
9.1	Must supply preassembled unit, ready to use.	



Biomedical Engineer
NEC No.- 83941